


Labeling Of Ultrasound Transmission Gel Bottles

When seal
present, use
Manufacturer's
expiration date



Once
opened,
write open
date with
permanent
Sharpie in
the
designated
space

Once bottle opened,
the expiration date is 30 days from open date

Write the following on the bottle without writing over product
information

O-11/01/2022 (Date opened)

E- 12/01/2022 (Expiration date- 30 days after opened)